

C E R T I F I C A T E
of Conformity



Registration No.: AK 50467699 0001

Report No.: 60372045 001

Holder: Shenzhen Everbest Machinery Industry Co., Ltd.
19th Building, 5th Region,
Baiwangxin Industrial Park,
Songbai Rd., Baimang, Xili, Nanshan
518108 Shenzhen
P.R. China

Product: Medical Device
Non-Contact Forehead Infrared Thermometer

Identification: Type Designation: DT-8807H, DT-8807S, DT-8807, DT-8817H
Trade Mark : CEM
Serial No. : N.A.
Remark: Refer to 60355772 001 and 60356703 001 for details.


Tested acc. to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12
IEC 60601-1-11:2015
EN 60601-1-11:2015
IEC 60601-1-2:2014
EN 60601-1-2:2015
ISO 80601-2-56:2017+A1

The certificate of conformity refers to the above mentioned product. This is to certify that the specimen is in conformity with the assessment requirement mentioned above. This certificate does not imply assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of conformity.

Date 13.05.2020



Certification Body


Dipl.-Ing. Univ. S. O. Steinke

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg